California Medical Device Recall Information

Recall Name
Baxter International Recalls INTRAVIA Containers Due to Particulate Matter

Recall Date	Product Description	Recalling Firm	Recall Reason
09/16/14	**INTRAVIA** Empty Containers with PVC Ports	**Baxter Healthcare Corporation** Deerfield, IL	*Due to complaints of particulate matter found inside the sterile fluid path.*
Recall Class	**Product Identification**	**Distribution**	**Affected Dates**
I	Two Lots: • INTRAVIA Container, 150 mL Capacity, Product Code: **2B8011** **Lot: UR13D15112** • INTRAVIA Container, Empty 500 mL Capacity, Product Code: **2B8013** **Lot: UR13K14095**	**CA**, nationwide	*Lot # UR13D15112:* distributed between April 26 and June 20, 2013; *Lot # UR13K14095:* distributed between November 27, 2013 and March 10, 2014.

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm423933.htm